

lifetime limited warranty

Your Drive brand product is warranted to be free of defects in materials and workmanship for the lifetime of the original consumer purchaser.

This device was built to exacting standards and carefully inspected prior to shipment. This Lifetime Limited Warranty is an expression of our confidence in the materials and workmanship of our products and our assurance to the consumer of years of dependable service.

In the event of a defect covered by this warranty, we will, at our option, repair or replace the device.

This warranty does not cover device failure due to owner misuse or negligence, or normal wear and tear. The warranty does not extend to non-durable components, such as rubber accessories, casters, and grips, which are subject to normal wear and need periodic replacement.

If you have a question about your Drive device or this warranty, please contact an authorized Drive dealer.

drive
MEDICAL DESIGN & MANUFACTURING

© 2005 Medical Depot, Inc. All rights reserved.
Drive is a trademark of Medical Depot, Inc.
Port Washington N.Y. 11050 USA Made in China

oversized drop arm commode

Distributor:
DeVilbiss Healthcare GmbH
Kamenzer Straße 3
68309 Mannheim
Germany
Tel: +49 (0) 621 17898-0
Fax: +49 (0) 621 17898-111
kontakt@devilbiss.com
www.drivedevilbiss-int.com

Manufacturer:
DriveDeVilbiss Healthcare Ltd
Heathfield Lane
Birkenshaw
West Yorkshire
BD11 2HW
Tel: +44 (0) 845 0600 333
Fax: +44 (0) 845 0600 334
info@drivedevilbiss.co.uk
www.drivedevilbiss.co.uk



www.drivemedical.com

drive
MEDICAL DESIGN & MANUFACTURING

commode assembly instructions and precautions

FIGURE 1



assembling commode seat and lid

1. Before assembling, allow toilet seat and lid to reach room temperature if exposed to cold. This will help prevent the seat and lid clamps from cracking while being assembled onto commode back frame rail.
- 2.. Align the toilet seat clamps onto the back frame rail. (Both clamps should sit outside the pail brackets.) Push down on seat until clamps snap firmly into place.
3. Slide pail in from front of commode.

FIGURE 2



adjusting leg height

1. Before use, adjust height of each leg to your specific needs, making sure all legs are at the same height. When desired height is reached, lock legs into place by aligning buttons into proper holes, per illustration above.
2. To be sure legs are locked into place by pushing down on or by pulling legs.

additional precautions

1. Toilet seat must be in down position before use. Failure to do so may result in breakage of commode and possible injury.
2. Do NOT install commode without first reading and understanding this assembly and instruction sheet. If you can not understand these instructions, contact a healthcare, dealer or technical professional before attempting to install. Otherwise, injury or damage may occur.
3. Inspect all parts for shipping damage before assembly. If there is shipping damage - DO NOT USE. Contact dealer/carrier for further instructions.
4. Make certain that the height adjusting snap buttons fully protrude through the same respective hole of each leg extension. This will ensure that the leg extensions are securely locked in position and an even height is achieved (see Figure 1).
5. Ensure ALL screws, nuts and/or bolts are tightened.
6. Check the rubber tips on the leg extensions for rips, wear or if they are missing. Immediately replace any or all if any of these imperfections exist.
7. All four leg extensions with rubber tips MUST touch the floor simultaneously at all times.
8. Users with limited physical strength should be supervised or assisted while using commode.